

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02209	Calib. Date:	12/02/2021
Certification File No.:	02210	Cert. Date:	12/02/2021
Linearity File No.:	02211	Lin. Date:	12/02/2021
Solution File No.:	02246	Soln. Date:	03/26/2022
Sequential File No.:	02246	File Date:	03/26/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-0110
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	1427

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:56S		
Control 1 EC	-.---%	08:56S	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	08:56S	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: _____

364

Badge No.: 364

Date: 03/26/2022

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Gregory Tenderenda
Passaic City

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142.04

THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C

A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

[Signature]
 /
 SHERIFF
 NEW JERSEY STATE POLICE

[Signature]
 /
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013- 2/14/2013

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2.	<u>9-11-20</u>	<u>MONTICELLO</u>	<u>[Signature]</u>
3.	<u>2-5-22</u>	<u>WAYNE</u>	<u>[Signature]</u>
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 293B (Rev. 01/18)