

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02259	Calib. Date:	04/20/2022
Certification File No.:	02260	Cert. No.:	00041
Linearity File No.:	02261	Cert. Date:	04/20/2022
Solution File No.:	02262	Lin. No.:	00032
Sequential File No.:	02262	Lin. Date:	04/20/2022
		Soln. No.:	00371
		Soln. Date:	04/20/2022
		File Date:	04/20/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-111
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	0911

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:17D		
Control 1 EC	0.099%	09:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:19D		
Control 2 EC	0.099%	09:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:20D		
Control 3 EC	0.100%	09:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.  
 Temperature Probe Serial Number: DDWJ P2 - 280 RW

**Changed By:**

Last Name: GRLICA

First Name: ROBERT

MI: M.

Signature: TRK [Signature] # 7034

Badge No.: 7034

Date: 04/20/2022

DEPARTMENT OF  
*Law and Public Safety*  
 This is to certify that  
**Gregory Tenderenda**

**Passaic City**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF  
 THE LAWS OF 1946 IN THE OPERATION OF THE Alcotest 7110 MKIII-C

A METHOD TO DETERMINE INTOXICATION  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

*[Signature]*  
 COLONEL  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013-2/14/2013

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2.	<u>9-11-20</u>	<u>MORRIS COTTA</u>	<i>[Signature]</i>
3.	<u>9-3-22</u>	<u>WAYNE</u>	<i>[Signature]</i>
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 293B (Rev. 01/10)