

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0031  
Location: PASSAIC CITY POLICE  
Calibration File No.: 02259 Calib. Date: 04/20/2022 Calib. No.: 00041  
Certification File No.: 02260 Cert. Date: 04/20/2022 Cert. No.: 00032  
Linearity File No.: 02261 Lin. Date: 04/20/2022 Lin. No.: 00032  
Solution File No.: 02304 Soln. Date: 06/19/2022 Soln. No.: 00375  
Sequential File No.: 02304 File Date: 06/19/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-111  
Control Solution %: 0.100% Expires: 09/01/2023  
Solution Control Lot: 21360 Bottle No.: 1061

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:06D		
Control 1 EC	0.098%	11:07D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:07D		
Control 2 EC	0.097%	11:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09D		
Control 3 EC	0.097%	11:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: \_\_\_\_\_



Badge No.: 364

Date: 06/19/2022

DEPARTMENT OF  
*Law and Public Safety*  
 This is to certify that  
**Gregory Tenderenda**

**Passaic City**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF  
 THE LAWS OF 1964 IN THE OPERATION OF THE: Alcotest 7110 MKIII-C  
 A METHOD TO DETERMINE INTOXICATION.  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

*[Signature]*  
 COLONEL  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013- 2/14/2013

DATE	Refresher Course PLACE	INSTRUCTOR
1. <u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2. <u>9-11-20</u>	<u>MORRIS COPA</u>	<i>[Signature]</i>
3. <u>2-3-22</u>	<u>WAYNE</u>	<i>[Signature]</i>
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____