

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02259	Calib. Date:	04/20/2022
Certification File No.:	02260	Cert. Date:	04/20/2022
Linearity File No.:	02261	Lin. Date:	04/20/2022
Solution File No.:	02313	Soln. Date:	06/29/2022
Sequential File No.:	02313	File Date:	06/29/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-111
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	1006

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:24D		
Control 1 EC	0.097%	10:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:25D		
Control 2 EC	0.097%	10:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:26D		
Control 3 EC	0.098%	10:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: P.O.  #364

Badge No.: 364

Date: 06/29/2022

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Gregory Tenderenda

Passaic City

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1964 IN THE OPERATION OF THE Alcotest 7110 MKIII-C

AS A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

[Signature]
 COLONEL
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013-2/14/2013

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2.	<u>9-11-20</u>	<u>Mt. Pleasant</u>	<i>[Signature]</i>
3.	<u>2-3-22</u>	<u>WAYNE</u>	<i>[Signature]</i>
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 293B (Rev. 01/10)