

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02330 Calib. Date: 08/24/2022 Calib. No.: 00042
Certification File No.: 02331 Cert. Date: 08/24/2022 Cert. No.: 00033
Linearity File No.: 02332 Lin. Date: 08/24/2022 Lin. No.: 00033
Solution File No.: 02388 Soln. Date: 12/25/2022 Soln. No.: 00390
Sequential File No.: 02388 File Date: 12/25/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-111
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 1491

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:42S		
Control 1 EC	0.096%	17:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	17:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:43S		
Control 2 EC	0.094%	17:44S	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.095%	17:44S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	17:44S		
Control 3 EC	0.094%	17:45S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.094%	17:45S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	17:46S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: _____



Badge No.: 364

Date: 12/25/2022

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Gregory Tenderenda
Passaic City

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C

A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

[Signature]
 FOREMAN
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013- 2/14/2013

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2.	<u>9-11-20</u>	<u>MARY SCOPA</u>	<i>[Signature]</i>
3.	<u>2-5-22</u>	<u>WAYNE</u>	<i>[Signature]</i>
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 293B (Rev. 01/18)