

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0031  
Location: PASSAIC CITY POLICE  
Calibration File No.: 02330 Calib. Date: 08/24/2022 Calib. No.: 00042  
Certification File No.: 02331 Cert. Date: 08/24/2022 Cert. No.: 00033  
Linearity File No.: 02332 Lin. Date: 08/24/2022 Lin. No.: 00033  
Solution File No.: 02387 Soln. Date: 12/25/2022 Soln. No.: 00389  
Sequential File No.: 02387 File Date: 12/25/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-111  
Control Solution %: 0.100% Expires: 05/09/2024  
Solution Control Lot: 22160 Bottle No.: 1134

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:59S		
Control 1 EC	0.098%	14:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:00S		
Control 2 EC	0.092%	15:01S	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.093%	15:01S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:02S		
Control 3 EC	0.096%	15:02S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.094%	15:02S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:03S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: \_\_\_\_\_



Badge No.: 364

Date: 12/25/2022

DEPARTMENT OF  
*Law and Public Safety*  
 This is to certify that

**Gregory Tenderenda**

**Passaic City**

IS QUALIFIED AND CAPABLE TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF  
 THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C

A METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

*[Signature]*  
 COLONEL  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013-2/14/2013

DATE	Refresher Course PLACE	INSTRUCTOR
1. <u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2. <u>9-11-20</u>	<u>MARYS COPA</u>	<i>[Signature]</i>
3. <u>2-5-22</u>	<u>WAYNE</u>	<i>[Signature]</i>
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____