

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02330	Calib. Date: 08/24/2022
Certification File No.:	02331	Calib. No.: 00042
Linearity File No.:	02332	Cert. Date: 08/24/2022
Solution File No.:	02392	Cert. No.: 00033
Sequential File No.:	02392	Lin. Date: 08/24/2022
		Lin. No.: 00033
		Soln. Date: 12/31/2022
		Soln. No.: 00392
		File Date: 12/31/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-111
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 1319

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:40S		
Control 1 EC	0.099%	10:40S	34.0°C	AMBIENT AIR CHECK ERR
Control 1 IR	-.---%	10:40S	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	10:41S		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA	First Name: GREGORY	MI: -
Signature: 	Badge No.: 364	Date: 12/31/2022

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Gregory Tenderenda
Passaic City

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C

A METHOD TO DETERMINE INTOXICATION,
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

[Signature]
 COLONEL
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013- 2/14/2013

DATE	Refresher Course PLACE	INSTRUCTOR
1. <u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2. <u>9-11-20</u>	<u>W.P.S. COPIA</u>	<i>[Signature]</i>
3. <u>2-3-22</u>	<u>WAYNE</u>	<i>[Signature]</i>
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

S.P. 293B (Rev. 01/19)