

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02209	Calib. Date: 12/02/2021
Certification File No.:	02210	Calib. No.: 00040
Linearity File No.:	02211	Cert. Date: 12/02/2021
Solution File No.:	02220	Cert. No.: 00031
Sequential File No.:	02220	Lin. Date: 12/02/2021
		Lin. No.: 00031
		Soln. Date: 01/04/2022
		Soln. No.: 00362
		File Date: 01/04/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-0110
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0825

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:13S		
Control 1 EC	0.096%	02:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	02:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:14S		
Control 2 EC	0.094%	02:15S	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.097%	02:15S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	02:16S		
Control 3 EC	0.094%	02:16S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.096%	02:16S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	02:17S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: _____




Badge No.: 364

Date: 01/04/2022


DEPARTMENT OF
Traffic and Public Safety
 This is to certify that
Gregory Tenderenda
Passaic City

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
 THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C
 A METHOD TO DETERMINE INTOXICATION.
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

 DANIEL J. KELLY
 CLERK
 NEW JERSEY STATE POLICE

 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013-2/14/2013

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2.	<u>9-11-20</u>	<u>MUSKOGEE</u>	
3.	<u>9-22-22</u>	<u>WAYNE</u>	
4.			
5.			
6.			
7.			
8.			
9.			