

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02209	Calib. Date: 12/02/2021
Certification File No.:	02210	Calib. No.: 00040
Linearity File No.:	02211	Cert. Date: 12/02/2021
Solution File No.:	02219	Cert. No.: 00031
Sequential File No.:	02219	Lin. Date: 12/02/2021
		Lin. No.: 00031
		Soln. Date: 01/04/2022
		Soln. No.: 00361
		File Date: 01/04/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-0110
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0298

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:55S		
Control 1 EC	0.099%	00:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:57S		
Control 2 EC	0.096%	00:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.095%	00:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:58S		
Control 3 EC	0.096%	00:59S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.093%	00:59S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	00:59S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: \_\_\_\_\_



Badge No.: 364

Date: 01/04/2022

DEPARTMENT OF  
*Law and Public Safety*  
 This is to certify that  
**Gregory Tenderenda**

**Passaic City**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

TITLE LAWS OF 1946 IN THE OPERATION OF THE Alcotest 7110 MKIII-C

A METHOD TO DETERMINE INTOXICATION,  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

*[Signature]*  
 COLONEL  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013-2/14/2013

DATE	Refresher Course PLACE	INSTRUCTOR
1. <u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2. <u>9-11-20</u>	<u>MARYSCOPA</u>	<i>[Signature]</i>
3. <u>8-5-22</u>	<u>WAYNE</u>	<i>[Signature]</i>
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 293B (Rev. 01/18)