

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02209 Calib. Date: 12/02/2021 Calib. No.: 00040
Certification File No.: 02210 Cert. Date: 12/02/2021 Cert. No.: 00031
Linearity File No.: 02211 Lin. Date: 12/02/2021 Lin. No.: 00031
Solution File No.: 02235 Soln. Date: 02/05/2022 Soln. No.: 00365
Sequential File No.: 02235 File Date: 02/05/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-0110
Control Solution %: 0.100% Expires: 09/01/2023
Solution Control Lot: 21360 Bottle No.: 0962

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:02S		
Control 1 EC	0.098%	07:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:04S		
Control 2 EC	0.097%	07:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	07:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:05S		
Control 3 EC	0.096%	07:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: _____



Badge No.: 364

Date: 02/05/2022

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Gregory Tenderenda
Passaic City

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATHALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1964 IN THE OPERATION OF THE Alcotest 7110 MKIII-C

A METHOD TO DETERMINE INTOXICATION,
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

[Signature]
 COLONEL
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013-2/14/2013

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2.	<u>9-11-20</u>	<u>MORRIS COPA</u>	<i>[Signature]</i>
3.	<u>9-3-22</u>	<u>WAYNE</u>	<i>[Signature]</i>
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 293B (Rev. 01/18)