

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02209	Calib. Date:	12/02/2021
Certification File No.:	02210	Cert. Date:	12/02/2021
Linearity File No.:	02211	Lin. Date:	12/02/2021
Solution File No.:	02247	Soln. Date:	03/26/2022
Sequential File No.:	02247	File Date:	03/26/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-0110
Solution Control Lot:	21360	Expires:	09/01/2023
		Bottle No.:	1275

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:30S		
Control 1 EC	0.099%	11:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31S		
Control 2 EC	0.098%	11:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33S		
Control 3 EC	0.098%	11:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:			
Last Name: TENDERENDA	First Name: GREGORY		MI: -
Signature: 	364	Badge No.: 364	Date: 03/26/2022

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Gregory Tenderenda
Passaic City

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
 THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C
 A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August
 TWO THOUSAND AND Twenty

[Signature]
 COUNSEL
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013- 2/14/2013

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2.	<u>9-11-20</u>	<u>Morris COPA</u>	<i>[Signature]</i>
3.	<u>2-5-22</u>	<u>WAYNE</u>	<i>[Signature]</i>
4.			
5.			
6.			
7.			
8.			
9.			