

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02330	Calib. Date: 08/24/2022
Certification File No.:	02331	Calib. No.: 00042
Linearity File No.:	02332	Cert. Date: 08/24/2022
Solution File No.:	02397	Cert. No.: 00033
Sequential File No.:	02397	Lin. Date: 08/24/2022
		Lin. No.: 00033
		Soln. Date: 01/01/2023
		Soln. No.: 00395
		File Date: 01/01/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-111
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 1320

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:11S		
Control 1 EC	0.099%	12:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13S		
Control 2 EC	0.097%	12:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:14S		
Control 3 EC	0.097%	12:15S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:16S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: *P.O. G. Tenderenda #364*

Badge No.: 364

Date: 01/01/2023

DEPARTMENT OF  
*Health and Public Safety*  
 This is to certify that

Gregory Tenderenda  
 Passaic City

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL DRUG ANALYSES PURSUANT TO CHAPTER 142C

THE LAWS OF THE STATE IN THE OPERATION OF THE Alcotest 7110 MKIII-C

A METHOD TO DETERMINE INTOXICATION:  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

*David J. Peltz*  
 CLERK  
 NEW JERSEY STATE POLICE

*Gregory Tenderenda*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013- 2/14/2013

DATE	Refresher Course PLACE	INSTRUCTOR
1. 7/31/2018	M.C.P.A.	Roushinko
2. 9-11-20	Marys COPA	<i>[Signature]</i>
3. 8-3-22	WAYNE	<i>[Signature]</i>
4.		
5.		
6.		
7.		
8.		
9.		