

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02407	Calib. Date:	02/07/2023
Certification File No.:	02408	Cert. No.:	00043
Linearity File No.:	02409	Cert. Date:	02/07/2023
Solution File No.:	02424	Lin. Date:	02/07/2023
Sequential File No.:	02424	Soln. Date:	03/14/2023
		File Date:	03/14/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-0110
Solution Control Lot:	22160	Expires:	05/09/2024
		Bottle No.:	1472

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:31S		
Control 1 EC	0.099%	08:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32S		
Control 2 EC	0.096%	08:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	08:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:34S		
Control 3 EC	0.096%	08:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.095%	08:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: PO [Handwritten Signature] #364

Badge No.: 364

Date: 03/14/2023

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Gregory Tenderenda
 Passaic City

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

TITLE 14 OF 1964 IN THE OPERATION OF THE Alcotest 7110 MKIII-C

A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

[Signature]
 / S. J. P. O. E.
 CLERK
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013-2/14/2013

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2.	<u>9-11-20</u>	<u>MONTICELLO</u>	<i>[Signature]</i>
3.	<u>2-3-22</u>	<u>WAYNE</u>	<i>[Signature]</i>
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 299B (Rev. 01/10)