

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0031 |
| Location: | PASSAIC CITY POLICE | |
| Calibration File No.: | 02407 | Calib. Date: 02/07/2023 |
| Certification File No.: | 02408 | Calib. No.: 00043 |
| Linearity File No.: | 02409 | Cert. Date: 02/07/2023 |
| Solution File No.: | 02441 | Cert. No.: 00034 |
| Sequential File No.: | 02441 | Lin. Date: 02/07/2023 |
| | | Lin. No.: 00034 |
| | | Soln. Date: 05/01/2023 |
| | | Soln. No.: 00403 |
| | | File Date: 05/01/2023 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDXB S3-0110 |
| Solution Control Lot: | 22160 | Expires: 05/09/2024 |
| | | Bottle No.: 0989 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 11:07D | | |
| Control 1 EC | 0.098% | 11:08D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 11:08D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:08D | | |
| Control 2 EC | 0.097% | 11:09D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 11:09D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:10D | | |
| Control 3 EC | 0.097% | 11:10D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 11:10D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:11D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: _____

Badge No.: 364

Date: 05/01/2023

DEPARTMENT OF
Water and Public Safety
 This is to certify that

Gregory Tenderenda
Passaic City

IS QUALIFIED AND COMMITTED TO CONDUCT CHEMICAL BAGATHANALYSES PURSUANT TO CHAPTER 146 OF THE LAWS OF 1964 IN THE OPERATION OF THE Alcotest 7110 MKIII-C AS A METHOD TO DETERMINE INTOXICATION:
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August
 TWO THOUSAND AND Twenty

[Signature]
 ATTORNEY GENERAL
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013-2/14/2013

| | DATE | Refresher Course PLACE | INSTRUCTOR |
|----|------------------|------------------------|--------------------|
| 1. | <u>7/31/2018</u> | <u>M.C.P.A.</u> | <u>Roushinko</u> |
| 2. | <u>9-11-20</u> | <u>MONTICELLO</u> | <u>[Signature]</u> |
| 3. | <u>2-3-22</u> | <u>WAYNE</u> | <u>[Signature]</u> |
| 4. | | | |
| 5. | | | |
| 6. | | | |
| 7. | | | |
| 8. | | | |
| 9. | | | |