

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02407	Calib. Date: 02/07/2023
Certification File No.:	02408	Calib. No.: 00043
Linearity File No.:	02409	Cert. Date: 02/07/2023
Solution File No.:	02437	Lin. Date: 02/07/2023
Sequential File No.:	02437	Lin. No.: 00034
		Soln. Date: 04/16/2023
		Soln. No.: 00402
		File Date: 04/16/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-0110
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 1473

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:11D		
Control 1 EC	0.097%	00:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:13D		
Control 2 EC	0.096%	00:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	00:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:14D		
Control 3 EC	0.096%	00:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	00:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: _____

364

Badge No.: 364

Date: 04/16/2023

DEPARTMENT OF
Justice and Public Safety
 This is to certify that

Gregory Tenderenda
Passaic City

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL DRUG ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C

A METHOD OF DOCUMENTATION:

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

Devin P. Kelly ASST
 CLERK ATTORNEY GENERAL
 NEW JERSEY STATE BAR

ORIGINAL COURSE DATES 2/11/2013-2/14/2013

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>7/31/2018</u>	<u>M.C.P.A.</u>	<u>Roushinko</u>
2.	<u>9-11-20</u>	<u>MARLBOROUGH</u>	<u>[Signature]</u>
3.	<u>8-3-12</u>	<u>WAYNE</u>	<u>[Signature]</u>
4.			
5.			
6.			
7.			
8.			
9.			

S.P. 2932 (Rev. 01/18)