

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02465	Calib. Date: 01/19/2024
Certification File No.:	02466	Calib. No.: 00046
Linearity File No.:	02467	Cert. Date: 01/19/2024
Solution File No.:	02474	Cert. No.: 00035
Sequential File No.:	02474	Lin. Date: 01/19/2024
		Lin. No.: 00035
		Soln. Date: 02/10/2024
		Soln. No.: 00414
		File Date: 02/10/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-111
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 1324

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:43S		
Control 1 EC	0.096%	10:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44S		
Control 2 EC	0.094%	10:45S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.096%	10:45S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:46S		
Control 3 EC	0.095%	10:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.095%	10:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: _____

P.O. [Signature] 364

Badge No.: 364

Date: 02/10/2024

DEPARTMENT OF
Water and Public Safety
 This is to certify that

Gregory Tenderenda

Passaic City

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF

THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C
 A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 3rd DAY OF August

TWO THOUSAND AND Twenty

[Signature]
 CLARENCE
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 2/11/2013-2/14/2013

DATE	Refresher Course PLACE	INSTRUCTOR
1. 7/31/2018	M.C.P.A.	Roushinko
2. 9-11-20	Mary K COPA	<i>[Signature]</i>
3. 2-3-12	WAYNE	<i>[Signature]</i>
4.		
5.		
6.		
7.		
8.		
9.		

S.P. 293B (Rev. 01/18)