

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02465	Calib. Date:	01/19/2024
Certification File No.:	02466	Cert. Date:	01/19/2024
Linearity File No.:	02467	Lin. Date:	01/19/2024
Solution File No.:	02481	Soln. Date:	03/04/2024
Sequential File No.:	02481	File Date:	03/04/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-111
Solution Control Lot:	22160	Expires:	05/09/2024
		Bottle No.:	1395

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:16S		
Control 1 EC	0.097%	16:16S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	16:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17S		
Control 2 EC	0.096%	16:18S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:19S		
Control 3 EC	0.096%	16:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	16:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: PO. [Handwritten Signature] '364

Badge No.: 364

Date: 03/04/2024