

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02465	Calib. Date: 01/19/2024
Certification File No.:	02466	Calib. No.: 00046
Linearity File No.:	02467	Cert. Date: 01/19/2024
Solution File No.:	02486	Cert. No.: 00035
Sequential File No.:	02486	Lin. Date: 01/19/2024
		Lin. No.: 00035
		Soln. Date: 03/22/2024
		Soln. No.: 00418
		File Date: 03/22/2024
Calibrating Unit:	4	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-111
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 1400

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:55S		
Control 1 EC	0.100%	12:56S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:56S		
Control 2 EC	0.099%	12:57S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:58S		
Control 3 EC	0.099%	12:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: \_\_\_\_\_

*P.O. [Signature] 364*

Badge No.: 364

Date: 03/22/2024