

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02465 Calib. Date: 01/19/2024 Calib. No.: 00046
Certification File No.: 02466 Cert. Date: 01/19/2024 Cert. No.: 00035
Linearity File No.: 02467 Lin. Date: 01/19/2024 Lin. No.: 00035
Solution File No.: 02491 Soln. Date: 04/01/2024 Soln. No.: 00420
Sequential File No.: 02491 File Date: 04/01/2024

Calibrating Unit: 4 Model No.: CU-34 Serial No.: DDXB S3-111
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 1317

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:49S		
Control 1 EC	0.099%	08:49S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50S		
Control 2 EC	0.098%	08:51S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:51S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:51S		
Control 3 EC	0.098%	08:52S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:53S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: P.O.  #364

Badge No.: 364

Date: 04/01/2024