

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02465 Calib. Date: 01/19/2024 Calib. No.: 00046
Certification File No.: 02466 Cert. Date: 01/19/2024 Cert. No.: 00035
Linearity File No.: 02467 Lin. Date: 01/19/2024 Lin. No.: 00035
Solution File No.: 02495 Soln. Date: 04/18/2024 Soln. No.: 00421
Sequential File No.: 02495 File Date: 04/18/2024

Calibrating Unit: 4 Model No.: CU-34 Serial No.: DDXB S3-111
Control Solution %: 0.100% Expires: 02/01/2025
Solution Control Lot: 23040 Bottle No.: 0579

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 10:51D | | |
| Control 1 EC | 0.098% | 10:52D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 10:52D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:52D | | |
| Control 2 EC | 0.098% | 10:53D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 10:53D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:54D | | |
| Control 3 EC | 0.098% | 10:54D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 10:54D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:55D | | |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: P.O.  364

Badge No.: 364

Date: 04/18/2024