

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02465	Calib. Date:	01/19/2024
Certification File No.:	02466	Cert. No.:	00046
Linearity File No.:	02467	Cert. Date:	01/19/2024
Solution File No.:	02497	Lin. Date:	01/19/2024
Sequential File No.:	02497	Soln. Date:	04/22/2024
		Soln. No.:	00035
		File Date:	04/22/2024
Calibrating Unit:	4	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-111
Solution Control Lot:	23040	Expires:	02/01/2025
		Bottle No.:	0139

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:20D		
Control 1 EC	0.097%	08:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	08:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:22D		
Control 2 EC	0.096%	08:22D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	08:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23D		
Control 3 EC	0.096%	08:24D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	08:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature:  364

Badge No.: 364

Date: 04/22/2024