

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02465	Calib. Date: 01/19/2024
Certification File No.:	02466	Calib. No.: 00046
Linearity File No.:	02467	Cert. Date: 01/19/2024
Solution File No.:	02499	Cert. No.: 00035
Sequential File No.:	02499	Lin. Date: 01/19/2024
		Lin. No.: 00035
		Soln. Date: 04/27/2024
		Soln. No.: 00423
		File Date: 04/27/2024

Calibrating Unit: 4	Model No.: CU-34	Serial No.: DDXB S3-111
Control Solution %: 0.100%		Expires: 02/01/2025
Solution Control Lot: 23040		Bottle No.: 0076

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:42D		
Control 1 EC	0.101%	16:42D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:43D		
Control 2 EC	0.099%	16:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:44D		
Control 3 EC	0.098%	16:45D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature:  364

Badge No.: 364

Date: 04/27/2024