

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02501	Calib. Date: 05/09/2024
Certification File No.:	02502	Calib. No.: 00047
Linearity File No.:	02503	Cert. Date: 05/09/2024
Solution File No.:	02508	Cert. No.: 00036
Sequential File No.:	02508	Lin. Date: 05/09/2024
		Lin. No.: 00036
		Soln. Date: 05/20/2024
		Soln. No.: 00425
		File Date: 05/20/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-111
Solution Control Lot:	22410	Expires: 10/26/2024
		Bottle No.: 1459

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:50D		
Control 1 EC	0.099%	14:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:51D		
Control 2 EC	0.099%	14:52D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:52D		
Control 3 EC	0.098%	14:53D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: PO. HILL 364

Badge No.: 364

Date: 05/20/2024