

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02501	Calib. Date:	05/09/2024
Certification File No.:	02502	Cert. Date:	05/09/2024
Linearity File No.:	02503	Lin. Date:	05/09/2024
Solution File No.:	02517	Soln. Date:	06/26/2024
Sequential File No.:	02517	File Date:	06/26/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-111
Solution Control Lot:	24180	Expires:	05/14/2026
		Bottle No.:	0362

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:45D		
Control 1 EC	0.103%	12:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:46D		
Control 2 EC	0.101%	12:47D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:47D		
Control 3 EC	0.101%	12:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: _____

P.O. H. T. U. 301

Badge No.: 364

Date: 06/26/2024