

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02501 Calib. Date: 05/09/2024 Calib. No.: 00047
Certification File No.: 02502 Cert. Date: 05/09/2024 Cert. No.: 00036
Linearity File No.: 02503 Lin. Date: 05/09/2024 Lin. No.: 00036
Solution File No.: 02525 Soln. Date: 07/23/2024 Soln. No.: 00428
Sequential File No.: 02525 File Date: 07/23/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-0110
Control Solution %: 0.100% Expires: 05/14/2026
Solution Control Lot: 24180 Bottle No.: 0244

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 18:34D | | |
| Control 1 EC | 0.098% | 18:35D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 18:35D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:35D | | |
| Control 2 EC | 0.097% | 18:36D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 18:36D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:37D | | |
| Control 3 EC | 0.097% | 18:37D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 18:37D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 18:38D | | |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: 

Badge No.: 364

Date: 07/23/2024