

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02501	Calib. Date: 05/09/2024
Certification File No.:	02502	Calib. No.: 00047
Linearity File No.:	02503	Cert. Date: 05/09/2024
Solution File No.:	02543	Cert. No.: 00036
Sequential File No.:	02543	Lin. Date: 05/09/2024
		Lin. No.: 00036
		Soln. Date: 09/09/2024
		Soln. No.: 00434
		File Date: 09/09/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-0110
Solution Control Lot:	24180	Expires: 05/14/2026
		Bottle No.: 0083

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:32D		
Control 1 EC	0.097%	10:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:33D		
Control 2 EC	0.096%	10:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:34D		
Control 3 EC	0.096%	10:35D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: 

Badge No.: 364

Date: 09/09/2024