

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02501	Calib. Date:	05/09/2024
Certification File No.:	02502	Cert. No.:	00047
Linearity File No.:	02503	Cert. Date:	05/09/2024
Solution File No.:	02548	Lin. Date:	05/09/2024
Sequential File No.:	02548	Soln. Date:	09/15/2024
		File Date:	09/15/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-0110
Solution Control Lot:	24180	Expires:	05/14/2026
		Bottle No.:	0259

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:15D		
Control 1 EC	0.095%	11:15D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	11:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:16D		
Control 2 EC	0.094%	11:17D	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.097%	11:17D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	11:17D		
Control 3 EC	0.094%	11:18D	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.097%	11:18D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	11:19D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: 

Badge No.: 364

Date: 09/15/2024