

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02501 Calib. Date: 05/09/2024 Calib. No.: 00047
Certification File No.: 02502 Cert. Date: 05/09/2024 Cert. No.: 00036
Linearity File No.: 02503 Lin. Date: 05/09/2024 Lin. No.: 00036
Solution File No.: 02549 Soln. Date: 09/16/2024 Soln. No.: 00437
Sequential File No.: 02549 File Date: 09/16/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-0110
Control Solution %: 0.100% Expires: 05/14/2026
Solution Control Lot: 24180 Bottle No.: 0156

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:22D		
Control 1 EC	0.097%	09:22D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:23D		
Control 2 EC	0.097%	09:24D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:24D		
Control 3 EC	0.097%	09:25D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: PO. [Signature] 364

Badge No.: 364

Date: 09/16/2024