

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02555	Calib. Date:	09/26/2024
Certification File No.:	02556	Cert. Date:	09/26/2024
Linearity File No.:	02557	Lin. Date:	09/26/2024
Solution File No.:	02560	Soln. Date:	09/30/2024
Sequential File No.:	02560	File Date:	09/30/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-0110
Solution Control Lot:	24180	Expires:	05/14/2026
		Bottle No.:	0495

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:20D		
Control 1 EC	0.097%	09:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:21D		
Control 2 EC	0.096%	09:21D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:22D		
Control 3 EC	0.096%	09:23D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: 

Badge No.: 364

Date: 09/30/2024