

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02555	Calib. Date: 09/26/2024
Certification File No.:	02556	Calib. No.: 00048
Linearity File No.:	02557	Cert. Date: 09/26/2024
Solution File No.:	02563	Cert. No.: 00037
Sequential File No.:	02563	Lin. Date: 09/26/2024
		Lin. No.: 00037
		Soln. Date: 10/07/2024
		Soln. No.: 00440
		File Date: 10/07/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-0110
Solution Control Lot:	24180	Expires: 05/14/2026
		Bottle No.: 0019

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:14D		
Control 1 EC	0.097%	10:14D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15D		
Control 2 EC	0.096%	10:15D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	10:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16D		
Control 3 EC	0.096%	10:17D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: _____

Badge No.: 364

Date: 10/07/2024