

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0031  
Location: PASSAIC CITY POLICE  
Calibration File No.: 02555 Calib. Date: 09/26/2024 Calib. No.: 00048  
Certification File No.: 02556 Cert. Date: 09/26/2024 Cert. No.: 00037  
Linearity File No.: 02557 Lin. Date: 09/26/2024 Lin. No.: 00037  
Solution File No.: 02566 Soln. Date: 10/12/2024 Soln. No.: 00441  
Sequential File No.: 02566 File Date: 10/12/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-0110  
Control Solution %: 0.100% Expires: 05/14/2026  
Solution Control Lot: 24180 Bottle No.: 0480

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:57D		
Control 1 EC	0.097%	10:58D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58D		
Control 2 EC	0.096%	10:59D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:00D		
Control 3 EC	0.097%	11:00D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:01D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature:  364

Badge No.: 364

Date: 10/12/2024