

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02555 Calib. Date: 09/26/2024 Calib. No.: 00048
Certification File No.: 02556 Cert. Date: 09/26/2024 Cert. No.: 00037
Linearity File No.: 02557 Lin. Date: 09/26/2024 Lin. No.: 00037
Solution File No.: 02576 Soln. Date: 11/10/2024 Soln. No.: 00443
Sequential File No.: 02576 File Date: 11/10/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-0110
Control Solution %: 0.100% Expires: 05/14/2026
Solution Control Lot: 24180 Bottle No.: 0298

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:56S		
Control 1 EC	0.099%	08:57S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	08:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58S		
Control 2 EC	0.097%	08:58S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:59S		
Control 3 EC	0.096%	09:00S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature:

 364

Badge No.: 364

Date: 11/10/2024