

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02555 Calib. Date: 09/26/2024 Calib. No.: 00048
Certification File No.: 02556 Cert. Date: 09/26/2024 Cert. No.: 00037
Linearity File No.: 02557 Lin. Date: 09/26/2024 Lin. No.: 00037
Solution File No.: 02586 Soln. Date: 11/24/2024 Soln. No.: 00447
Sequential File No.: 02586 File Date: 11/24/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-0110
Control Solution %: 0.100% Expires: 05/14/2026
Solution Control Lot: 24180 Bottle No.: 0412

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 12:26S | | |
| Control 1 EC | 0.093% | 12:26S | 33.9°C | CONTROL TEST FAILED |
| Control 1 IR | 0.095% | 12:26S | 33.9°C | CONTROL TEST FAILED |
| Ambient Air Blank | 0.000% | 12:27S | | |
| Control 2 EC | 0.092% | 12:28S | 33.9°C | CONTROL TEST FAILED |
| Control 2 IR | 0.096% | 12:28S | 33.9°C | CONTROL TEST FAILED |
| Ambient Air Blank | 0.000% | 12:29S | | |
| Control 3 EC | 0.092% | 12:29S | 33.9°C | CONTROL TEST FAILED |
| Control 3 IR | 0.095% | 12:29S | 33.9°C | CONTROL TEST FAILED |
| Ambient Air Blank | 0.000% | 12:30S | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: 

Badge No.: 364

Date: 11/24/2024