

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02555 Calib. Date: 09/26/2024 Calib. No.: 00048
Certification File No.: 02556 Cert. Date: 09/26/2024 Cert. No.: 00037
Linearity File No.: 02557 Lin. Date: 09/26/2024 Lin. No.: 00037
Solution File No.: 02587 Soln. Date: 11/24/2024 Soln. No.: 00448
Sequential File No.: 02587 File Date: 11/24/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-0110
Control Solution %: 0.100% Expires: 05/14/2026
Solution Control Lot: 24180 Bottle No.: 0517

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:58S		
Control 1 EC	0.095%	13:59S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	13:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:00S		
Control 2 EC	0.094%	14:00S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.097%	14:00S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:01S		
Control 3 EC	0.094%	14:02S	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.097%	14:02S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:02S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: PO G. T. 364

Badge No.: 364

Date: 11/24/2024