

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02555 Calib. Date: 09/26/2024 Calib. No.: 00048
Certification File No.: 02556 Cert. Date: 09/26/2024 Cert. No.: 00037
Linearity File No.: 02557 Lin. Date: 09/26/2024 Lin. No.: 00037
Solution File No.: 02589 Soln. Date: 11/26/2024 Soln. No.: 00450
Sequential File No.: 02589 File Date: 11/26/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-0110
Control Solution %: 0.100% Expires: 05/14/2026
Solution Control Lot: 24180 Bottle No.: 0086

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:44S		
Control 1 EC	0.092%	17:45S	33.8°C	CONTROL TEST FAILED
Control 1 IR	0.092%	17:45S	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	17:45S		
Control 2 EC	0.091%	17:46S	33.8°C	CONTROL TEST FAILED
Control 2 IR	0.093%	17:46S	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	17:47S		
Control 3 EC	0.089%	17:48S	33.8°C	CONTROL TEST FAILED
Control 3 IR	0.091%	17:48S	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	17:48S		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: 

Badge No.: 364

Date: 11/26/2024