

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0031  
Location: PASSAIC CITY POLICE  
Calibration File No.: 02555 Calib. Date: 09/26/2024 Calib. No.: 00048  
Certification File No.: 02556 Cert. Date: 09/26/2024 Cert. No.: 00037  
Linearity File No.: 02557 Lin. Date: 09/26/2024 Lin. No.: 00037  
Solution File No.: 02590 Soln. Date: 11/26/2024 Soln. No.: 00451  
Sequential File No.: 02590 File Date: 11/26/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-0110  
Control Solution %: 0.100% Expires: 05/14/2026  
Solution Control Lot: 0392 Bottle No.: 0392

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:02S		
Control 1 EC	0.094%	19:03S	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.094%	19:03S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	19:03S		
Control 2 EC	0.093%	19:04S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.096%	19:04S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	19:05S		
Control 3 EC	0.092%	19:06S	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.095%	19:06S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	19:06S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: PO. [Handwritten Signature] 364

Badge No.: 24180

Date: 11/26/2024