

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02555	Calib. Date:	09/26/2024
Certification File No.:	02556	Cert. Date:	09/26/2024
Linearity File No.:	02557	Lin. Date:	09/26/2024
Solution File No.:	02592	Soln. Date:	12/04/2024
Sequential File No.:	02592	File Date:	12/04/2024
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-0110
Solution Control Lot:	24180	Expires:	05/14/2026
		Bottle No.:	0470

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:41S		
Control 1 EC	0.095%	20:42S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.095%	20:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:43S		
Control 2 EC	0.094%	20:43S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.095%	20:43S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	20:44S		
Control 3 EC	0.094%	20:45S	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.095%	20:45S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	20:46S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature:



Badge No.: 364

Date: 12/04/2024