

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02555	Calib. Date: 09/26/2024
Certification File No.:	02556	Calib. No.: 00048
Linearity File No.:	02557	Cert. Date: 09/26/2024
Solution File No.:	02593	Cert. No.: 00037
Sequential File No.:	02593	Lin. Date: 09/26/2024
		Lin. No.: 00037
		Soln. Date: 12/05/2024
		Soln. No.: 00454
		File Date: 12/05/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-0110
Solution Control Lot:	24180	Expires: 05/14/2026
		Bottle No.: 0364

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:57S		
Control 1 EC	0.095%	09:57S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.095%	09:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58S		
Control 2 EC	0.094%	09:59S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.095%	09:59S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:00S		
Control 3 EC	0.094%	10:00S	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.095%	10:00S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:01S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA	First Name: GREGORY	MI: -
Signature: <u>PO [Handwritten Signature] 364</u>	Badge No.: 364	Date: 12/05/2024