

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0031  
Location: PASSAIC CITY POLICE  
Calibration File No.: 02555 Calib. Date: 09/26/2024 Calib. No.: 00048  
Certification File No.: 02556 Cert. Date: 09/26/2024 Cert. No.: 00037  
Linearity File No.: 02557 Lin. Date: 09/26/2024 Lin. No.: 00037  
Solution File No.: 02595 Soln. Date: 12/09/2024 Soln. No.: 00456  
Sequential File No.: 02595 File Date: 12/09/2024

Calibrating Unit: WET Model No.: CU-34 Serial No.: ]  
Control Solution %: 0.100% Expires: 05/14/2026  
Solution Control Lot: 24180 Bottle No.: 0201

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:24S		
Control 1 EC	0.091%	12:24S	33.8°C	CONTROL TEST FAILED
Control 1 IR	0.091%	12:24S	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	12:25S		
Control 2 EC	0.088%	12:26S	33.8°C	CONTROL TEST FAILED
Control 2 IR	0.089%	12:26S	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	12:27S		
Control 3 EC	0.088%	12:28S	33.8°C	CONTROL TEST FAILED
Control 3 IR	0.091%	12:28S	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	12:28S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: PO. [Signature] '361

Badge No.: 364

Date: 12/09/2024