

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02555	Calib. Date: 09/26/2024
Certification File No.:	02556	Calib. No.: 00048
Linearity File No.:	02557	Cert. Date: 09/26/2024
Solution File No.:	02597	Cert. No.: 00037
Sequential File No.:	02597	Lin. Date: 09/26/2024
		Lin. No.: 00037
		Soln. Date: 12/16/2024
		Soln. No.: 00458
		File Date: 12/16/2024
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-0110
Solution Control Lot:	24180	Expires: 05/14/2026
		Bottle No.: 0380

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:34S		
Control 1 EC	0.087%	14:34S	33.8°C	CONTROL TEST FAILED
Control 1 IR	0.089%	14:34S	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:35S		
Control 2 EC	0.087%	14:36S	33.8°C	CONTROL TEST FAILED
Control 2 IR	0.089%	14:36S	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:37S		
Control 3 EC	0.086%	14:37S	33.8°C	CONTROL TEST FAILED
Control 3 IR	0.088%	14:37S	33.8°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:38S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: 

Badge No.: 364

Date: 12/16/2024