

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02598	Calib. Date: 12/18/2024
Certification File No.:	02599	Calib. No.: 00049
Linearity File No.:	02600	Cert. Date: 12/18/2024
Solution File No.:	02603	Cert. No.: 00038
Sequential File No.:	02603	Lin. Date: 12/18/2024
		Lin. No.: 00038
		Soln. Date: 01/08/2025
		Soln. No.: 00460
		File Date: 01/08/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-111
Solution Control Lot:	24180	Expires: 05/14/2026
		Bottle No.: 0363

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:57S		
Control 1 EC	0.099%	14:58S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:59S		
Control 2 EC	0.094%	14:59S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.094%	14:59S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:00S		
Control 3 EC	0.096%	15:01S	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.092%	15:01S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:02S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: _____

TO. TENDERENDA 1364

Badge No.: 364

Date: 01/08/2025