Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE				
Calibration File No.:	02598	Calib. Date:	12/18/2024	Calib. No.:	00049
Certification File No.:	02599	Cert. Date:	12/18/2024	Cert. No.:	00038
Linearity File No.:	02600	Lin. Date:	12/18/2024	Lin. No.:	00038
Solution File No.:	02604	Soln. Date:	01/19/2025	Soln. No.:	00461
Sequential File No.:	02604	File Date:	01/19/2025		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXB S3-111
Control Solution %:	0.100%			Expires:	05/14/2026
Solution Control Lot:	24180			Bottle No.:	0023

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:47S		
Control 1 EC	0.094%	00:47S	33.9°C	CONTROL TEST FAILED
Control 1 IR	0.095%	00:47S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	00:48S		
Control 2 EC	0.094%	00:49S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.095%	00:49S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	00:50S		
Control 3 EC	0.094%	00:50S	33.9°C	CONTROL TEST FAILED
Control 3 IR	0.095%	00:50S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	00:51S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA First Name: GREGORY MI: -

Signature: B. T #364 Badge No.: 364

Date: 01/19/2025