

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02598 Calib. Date: 12/18/2024 Calib. No.: 00049
Certification File No.: 02599 Cert. Date: 12/18/2024 Cert. No.: 00038
Linearity File No.: 02600 Lin. Date: 12/18/2024 Lin. No.: 00038
Solution File No.: 02605 Soln. Date: 01/30/2025 Soln. No.: 00462
Sequential File No.: 02605 File Date: 01/30/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-111
Control Solution %: 0.100% Expires: 07/23/2026
Solution Control Lot: 24260 Bottle No.: 1099

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:43S		
Control 1 EC	0.099%	13:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:45S		
Control 2 EC	0.099%	13:46S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46S		
Control 3 EC	0.099%	13:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature:  #364

Badge No.: 364

Date: 01/30/2025