

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02598	Calib. Date: 12/18/2024
Certification File No.:	02599	Calib. No.: 00049
Linearity File No.:	02600	Cert. Date: 12/18/2024
Solution File No.:	02608	Cert. No.: 00038
Sequential File No.:	02608	Lin. Date: 12/18/2024
		Lin. No.: 00038
		Soln. Date: 02/14/2025
		Soln. No.: 00463
		File Date: 02/14/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-111
Solution Control Lot:	24260	Expires: 07/23/2026
		Bottle No.: 0694

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:53S		
Control 1 EC	0.099%	17:54S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:55S		
Control 2 EC	0.097%	17:56S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	17:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:57S		
Control 3 EC	0.097%	17:57S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	17:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:58S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: 

Badge No.: 364

Date: 02/14/2025

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
 Location: PASSAIC CITY POLICE
 Calibration File No.: 02598
 Certification File No.: 02599
 Linearity File No.: 02600
 Solution File No.: 02605
 Sequential File No.: 02605

Serial No.: ARXA-0031

Calib. Date: 12/18/2024 Calib. No.: 00049
 Cert. Date: 12/18/2024 Cert. No.: 00038
 Lin. Date: 12/18/2024 Lin. No.: 00038
 Soln. Date: 01/30/2025 Soln. No.: 00462
 File Date: 01/30/2025

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 24260

Model No.: CU-34

Serial No.: DDXB S3-111
 Expires: 07/23/2026
 Bottle No.: 1099

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:43S		
Control 1 EC	0.099%	13:44S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:44S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:45S		
Control 2 EC	0.099%	13:46S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46S		
Control 3 EC	0.099%	13:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: T.O. Hill #304

Badge No.: 364

Date: 01/30/2025