

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0031
Location: PASSAIC CITY POLICE
Calibration File No.: 02598 Calib. Date: 12/18/2024 Calib. No.: 00049
Certification File No.: 02599 Cert. Date: 12/18/2024 Cert. No.: 00038
Linearity File No.: 02600 Lin. Date: 12/18/2024 Lin. No.: 00038
Solution File No.: 02616 Soln. Date: 03/17/2025 Soln. No.: 00465
Sequential File No.: 02616 File Date: 03/17/2025

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-111
Control Solution %: 0.100% Expires: 07/23/2026
Solution Control Lot: 24260 Bottle No.: 0958

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:06S		
Control 1 EC	0.099%	07:06S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:07S		
Control 2 EC	0.097%	07:08S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:09S		
Control 3 EC	0.097%	07:09S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: 

364

Badge No.: 364

Date: 03/17/2025