

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0031 |
| Location: | PASSAIC CITY POLICE | |
| Calibration File No.: | 02598 | Calib. Date: 12/18/2024 |
| Certification File No.: | 02599 | Calib. No.: 00049 |
| Linearity File No.: | 02600 | Cert. Date: 12/18/2024 |
| Solution File No.: | 02619 | Cert. No.: 00038 |
| Sequential File No.: | 02619 | Lin. Date: 12/18/2024 |
| | | Lin. No.: 00038 |
| | | Soln. Date: 03/28/2025 |
| | | Soln. No.: 00466 |
| | | File Date: 03/28/2025 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDXB S3-111 |
| Solution Control Lot: | 24260 | Expires: 07/23/2026 |
| | | Bottle No.: 0798 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 12:41S | | |
| Control 1 EC | 0.097% | 12:41S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.096% | 12:41S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:42S | | |
| Control 2 EC | 0.096% | 12:43S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 12:43S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:44S | | |
| Control 3 EC | 0.097% | 12:44S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 12:44S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:45S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: 

Badge No.: 364

Date: 03/28/2025