

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02598	Calib. Date: 12/18/2024
Certification File No.:	02599	Cert. No.: 00049
Linearity File No.:	02600	Cert. Date: 12/18/2024
Solution File No.:	02622	Lin. Date: 12/18/2024
Sequential File No.:	02622	Lin. No.: 00038
		Soln. Date: 04/08/2025
		Soln. No.: 00467
		File Date: 04/08/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-111
Solution Control Lot:	24260	Expires: 07/23/2026
		Bottle No.: 0810

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:59D		
Control 1 EC	0.099%	10:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		
Control 2 EC	0.098%	10:01D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		
Control 3 EC	0.098%	10:03D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: \_\_\_\_\_

*T. O. Tenderenda* 364

Badge No.: 364

Date: 04/08/2025