

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|-------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARXA-0031 |
| Location: | PASSAIC CITY POLICE | | |
| Calibration File No.: | 02598 | Calib. Date: | 12/18/2024 |
| Certification File No.: | 02599 | Cert. Date: | 12/18/2024 |
| Linearity File No.: | 02600 | Lin. Date: | 12/18/2024 |
| Solution File No.: | 02628 | Soln. Date: | 04/24/2025 |
| Sequential File No.: | 02628 | File Date: | 04/24/2025 |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDXB S3-111 |
| Solution Control Lot: | 24260 | Expires: | 07/23/2026 |
| | | Bottle No.: | 0780 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 10:27D | | |
| Control 1 EC | 0.098% | 10:27D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 10:27D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:28D | | |
| Control 2 EC | 0.097% | 10:29D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 10:29D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:30D | | |
| Control 3 EC | 0.098% | 10:30D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 10:30D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:31D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: PO. TENDERENDA 361

Badge No.: 364

Date: 04/24/2025