

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0031
Location:	PASSAIC CITY POLICE		
Calibration File No.:	02598	Calib. Date:	12/18/2024
Certification File No.:	02599	Cert. Date:	12/18/2024
Linearity File No.:	02600	Lin. Date:	12/18/2024
Solution File No.:	02638	Soln. Date:	05/12/2025
Sequential File No.:	02638	File Date:	05/12/2025
Calibrating Unit: WET		Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-111
Solution Control Lot:	24260	Expires:	07/23/2026
		Bottle No.:	0970

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:06D		
Control 1 EC	0.099%	10:07D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:07D		
Control 2 EC	0.098%	10:08D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09D		
Control 3 EC	0.099%	10:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TENDERENDA

First Name: GREGORY

MI: -

Signature: PO. G. T. M. 364

Badge No.: 364

Date: 05/12/2025