

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0031
Location:	PASSAIC CITY POLICE	
Calibration File No.:	02640	Calib. Date: 05/27/2025
Certification File No.:	02641	Calib. No.: 00050
Linearity File No.:	02642	Cert. Date: 05/27/2025
Solution File No.:	02643	Cert. No.: 00039
Sequential File No.:	02643	Lin. Date: 05/27/2025
		Lin. No.: 00039
		Soln. Date: 05/27/2025
		Soln. No.: 00470
		File Date: 05/27/2025
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-111
Solution Control Lot:	24260	Expires: 07/23/2026
		Bottle No.: 0957

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:01D		
Control 1 EC	0.100%	17:01D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:02D		
Control 2 EC	0.100%	17:03D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:03D		
Control 3 EC	0.098%	17:04D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	17:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJ P2-280 F.H.

Changed By:

Last Name: HURTADO	First Name: FRANCISCO	MI: A.
Signature: <u></u>	Badge No.: 7635	Date: 05/27/2025